



Current Data Parameters  
 NAME sample42  
 EXPNO 6  
 PROCNO 1

F2 - Acquisition Parameters  
 Date\_ 20010110  
 Time 18.03

INSTRUM dpx400  
 PROBHD 5 mm BBO Z-Gra  
 PULPROG invieagssi  
 TD 2048  
 SOLVENT CDCl3  
 NS 4  
 DS 8  
 SWH 2003.205 Hz  
 FIDRES 0.978127 Hz  
 AQ 0.5111808 sec  
 RG 3649.1  
 DW 249.600 usec  
 DE 6.00 usec  
 TE 300.0 K  
 CNSTO 1.0000000  
 D0 0.00000300 sec  
 D1 1.00000000 sec  
 D4 0.00172500 sec  
 D11 0.03000000 sec  
 D13 0.00000300 sec  
 D16 0.00030000 sec  
 d20 0.00137640 sec  
 d21 0.00130700 sec  
 D24 0.00172414 sec  
 INO 0.00004141 sec  
 l3 0

==== CHANNEL f1 =====  
 NUC1 1H  
 P1 8.10 usec  
 P2 20.40 usec  
 P28 2000.00 usec  
 PL1 -3.00 dB  
 SFO1 400.1322007 MHz

==== CHANNEL f2 =====  
 CPDPRG[2] garp  
 NUC2 13C  
 P3 7.00 usec  
 P4 16.00 usec  
 PCPD2 43.20 usec  
 PL2 -3.00 dB  
 PL12 12.00 dB  
 SFO2 100.6237964 MHz

==== GRADIENT CHANNEL =====  
 P16 1000.00 usec

F1 - Acquisition parameters  
 TD 256  
 SFO1 100.6238 MHz  
 FIDRES 94.325386 Hz  
 SW 119.988 ppm  
 FnmODE undefined

F2 - Processing parameters  
 SI 1024  
 SF 400.130094 MHz  
 WDW QSINE  
 SSB 2  
 LB 0 Hz  
 GB 0  
 PC 1.00

F1 - Processing parameters  
 SI 1024  
 MC2 echo-antiecho  
 SF 100.6127290 MHz  
 WDW QSINE  
 SSB 2  
 LB 0 Hz  
 GB 0