



```

Current Data Parameters
NAME      Un_69
EXPNO    4
PROCNO   1

F2 - Acquisition Parameters
Date_    20120313
Time     10.36
INSTRUM  robinson
PROBHD   5 mm PABBO BB-
PULPROG  hsqcedetgp
TD       1024
SOLVENT  DMSO
NS       2
DS       16
SWH      801.282 Hz
FIDRES   0.782502 Hz
AQ       0.6389760 sec
RG       18390.4
DW       624.000 usec
DE       6.50 usec
TE       299.0 K
CNST2    145.0000000
D0       0.00000300 sec
D1       2.00000000 sec
D4       0.00172414 sec
D11      0.03000000 sec
D13      0.00000400 sec
D16      0.00020000 sec
D21      0.00345000 sec
IN0      0.00006210 sec
ZGOPTNS

===== CHANNEL f1 =====
NUC1     1H
P1       10.90 usec
P2       21.80 usec
P28      1000.00 usec
PL1      0 dB
SFO1     400.1433212 MHz

===== CHANNEL f2 =====
CPDPRG[2] garp
NUC2     13C
P3       8.05 usec
P4       16.10 usec
PCPD2    65.00 usec
PL2      -2.00 dB
PL12     14.70 dB
SFO2     100.6293691 MHz

===== GRADIENT CHANNEL =====
GFNAM[1] SINE.100
GFNAM[2] SINE.100
GPZ1     80.00 %
GPZ2     20.10 %
P16      1000.00 usec

F1 - Acquisition parameters
TD       256
SFO1     100.6294 MHz
FIDRES   62.893356 Hz
SW       80.000 ppm
FnMODE   Echo-Antiecho

F2 - Processing parameters
SI       2048
SF       400.1400000 MHz
WDW      QSINE
SSB      2
LB       0 Hz
GB       0
PC       1.40

F1 - Processing parameters
SI       2048
MC2      echo-antiecho
SF       100.6152830 MHz
WDW      QSINE
SSB      2
LB       0 Hz
GB       0
    
```